



'Ready To Use Reagents'

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106
Lot Number 8194

Date 29.09.2008


Expiry Date 11.09.2009

Reagent	Lot Number
Phos 1	8194
Phos 2	8108
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 09 08	0.148
1	P20 09 08	0.289
1.5	P20 09 08	0.444
2	P20 09 08	0.572

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.287	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility

Dr. T. Russell