



'Ready To Use Reagents'

ver 2006:02

Certificate of Performance

Ammonia 2 Reagent

Product Code KR 109
Lot Number 9034

Date 03.02.2009


Expiry Date 30.01.2010

Reagent	Lot Number
AMM 1	9033
Amm 2	9034
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.021
0.25	AMM 10 09 08	0.226
0.50	AMM 10 09 08	0.432
0.77	AMM 10 09 08	0.654
1.00	AMM 10 09 08	0.848

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0198	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.8261	0.750 to 0.890	Pass
Regression	1.0000	0.990 to 1.0001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 109
 * Based on results obtained in our Research Facility

Tom Russell.
 Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.
Telephone 061 440207 Facsimile 061 440237