



'Ready To Use Reagents'

Certificate of Analysis

Nitrite as N

Product Code **KS 204**

Lot Number **9042**

Expiry Date 25.01.2010

Standard	Lot Number
KS 204	9042

Test	Result	Specification	Pass / Fail
Result ppm (as N)	999	990 to 1010	Pass

Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.

Telephone 061 440207

Facsimile 061 440237