



'Ready To Use Reagents'

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107
Lot Number 9095

Date 05.05.2009


Expiry Date 01.05.2010

Reagent	Lot Number
Phos 1	9039
Phos 2	9095
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 01 09	0.146
1	P20 01 09	0.293
1.5	P20 01 09	0.453
2	P20 01 09	0.585

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.001	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.295	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 107
 * Based on results obtained in our Research Facility

Tom Russell.

 Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.
 Telephone 061 440207 Facsimile 061 440237