Precision®

'Ready To Use Reagents'

Certificate of Analysis

Chloride as Cl

Product Code **KS 202**

Lot Number **9123**

Expiry Date 07.06.2010

Standard	Lot Number
KS 202	9123

Test	Result	Specification			Pass / Fail
Result ppm (as Cl)	1001	990	to	1010	Pass

Dr. T. Russell

Serosep laboratory diagnostics

An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.

Telephone 061 440207 Facsimile 061 440237