



'Ready To Use Reagents'

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106
Lot Number 9145

Date 17.08.2009


Expiry Date 17.08.2010

Reagent	Lot Number
Phos 1	9145
Phos 2	9095
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 07 09	0.150
1	P20 07 09	0.299
1.5	P20 07 09	0.459
2	P20 07 09	0.598

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.001	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.300	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility

Tom Russell.

 Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.

Telephone 061 440207

Facsimile 061 440237