



'Ready To Use Reagents'

Certificate of Analysis

Nitrate as NO₃

Product Code **KS 205**

Lot Number **9153**

Expiry Date 24.07.2010

Standard	Lot Number
KS 205	9153

Test	Result	Specification	Pass / Fail
Result ppm (as NO ₃)	1002	990 to 1010	Pass

Tom Russell.

Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.

Telephone 061 440207

Facsimile 061 440237