



*'Ready To Use Reagents'*

**Certificate of Performance**

**Phosphate 1 (Phos 1) Reagent**

**Product Code** KR106  
**Lot Number** 0198

**Date** 14.10.2010


**Expiry Date** 12.10.2011

Reagent	Lot Number
Phos 1	0198
Phos 2	9228
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 09 10	0.154
1	P20 09 10	0.295
1.5	P20 09 10	0.446
2	P20 09 10	0.587

\*serial dilution

**Calibration Curve**



	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.293	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

**Notes** Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 106  
 \* Based on results obtained in our Research Facility

*Tom Russell.*  
 \_\_\_\_\_  
 Dr. T. Russell



An ISO 9001:2000 Registered Company

Unit F5, Eastway Business Park, Ballysimon Road, Limerick, Ireland.

Telephone 061 440207

Facsimile 061 440237