



'Ready To Use Reagents'

Certificate of Analysis

Ammonia as NH₄

Product Code **KS 201**

Lot Number **0209**

Expiry Date 30.06.2011

Standard	Lot Number
KS 201	0209

Test	Result	Specification	Pass / Fail
Result ppm (as NH ₄)	1003	990 to 1010	Pass

Tom Russell.

Dr. T. Russell

