



'Ready To Use Reagents'

Certificate of Analysis

Chloride as Cl

Product Code **KS 202**

Lot Number **O210**

Expiry Date 30.09.2011

Standard	Lot Number
KS 202	O210

Test	Result	Specification	Pass / Fail
Result ppm (as Cl)	998	990 to 1010	Pass

Tom Russell.

Dr. T. Russell

