



*'Ready To Use Reagents'*

### Certificate of Analysis

#### Ammonia as N

**Product Code**            **KS 209**

**Lot Number**              **0215**

**Expiry Date**              31.12.2011

Standard	Lot Number
KS 209	0215

Test	Result	Specification	Pass / Fail
Result ppm (as N)	<b>993</b>	990            to            1010	<b>Pass</b>

*Tom Russell.*  

---

*Dr. T. Russell*

