



'Ready To Use Reagents'

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106
Lot Number 1071

Date 22.03.2011


Expiry Date 18.03.2012

Reagent	Lot Number
Phos 1	1071
Phos 2	0195
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.004
0.5	P20 03 11	0.153
1	P20 03 11	0.304
1.5	P20 03 11	0.464
2	P20 03 11	0.597

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.299	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility

Tom Russell.

Dr. T. Russell



An ISO 9001:2000 Registered Company

Annacotty Business Park, Annacotty, Limerick, Ireland.

Telephone 061 358190

Facsimile 061 358191