

Certificate of Performance

Ammonia 1 Reagent

Product Code **KR108**
Lot Number **1075**

Date 24.03.2011


Expiry Date 04.02.2012

Reagent	Lot Number
AMM 1	1075
Amm 2	1014
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.026
0.25	AMM 10 02 11	0.213
0.50	AMM 10 02 11	0.413
0.77	AMM 10 02 11	0.624
1.00	AMM 10 02 11	0.816

**serial dilution*

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0203	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.7899	0.750 to 0.890	Pass
Regression	0.9997	0.990 to 1.0001	Pass

Notes *Store between 2 & 8 °c*

**Stable on Board 1 day*

Barcode ID = 108

** Based on results obtained in our Research Facility*



Dr. T. Russell