Precision®

*Ready To Use Reagents'* Version 2011.01

# Certificate of Performance

## TON 2

Product Code KR104

**Date** 25.03.2011 **Expiry Date** 11.03.2012

| Reagent | Lot Number |
|---|---|
| TON 1 | 1033 |
| TON 2 | 1077 |
| TON 3 | 1027 |
| Test Method | LMTP03 |
| Test Instrument | Konelab 20 |
| Serial Number | P4919028 |

| Std. Value | Lot # | Abs. |
|---|---|---|
| 0.00 | | 0.007 |
| 2.50 | TON 100 03 11 | 0.194 |
| 5.00 | TON 100 03 11 | 0.364 |
| 7.69 | TON 100 03 11 | 0.531 |
| 10.00 | TON 100 03 11 | 0.658 |

*serial dilution

### Calibration Curve




| Test | | Result | Specification | | | Pass / Fail |
|---|---|---|---|---|---|---|
| Intercept | | 0.007 | -0.005 | to | 0.01 | Pass |
| Sensitivity | Abs/ppm | | | | | |
| | *0 - 2.50 mg/l* | 0.0748 | 0.0550 | to | 0.0800 | Pass |
| | *2.50 - 5.00 mg/l* | 0.0680 | 0.0600 | to | 0.0800 | Pass |
| | *5.00 - 7.69 mg/l* | 0.0621 | 0.0550 | to | 0.0800 | Pass |
| | *7.69 - 10.0 mg/l* | 0.0550 | 0.0500 | to | 0.0700 | Pass |

*Notes* *Store between 2 & 8 °C*

**Stable on Board TON 1 - 1 day*

** TON 2 & TON 3 - 5 days*

*Keep Tightly Sealed when not in use.*

*Do not return unused portions to container*

*Barcode ID- TON 1 103; TON 2 104; TON 3 105*

** Based on results obtained in our Research Facility*

Tom Russell.

*Dr. T. Russell*


Serosep laboratory diagnostics

**An ISO 9001:2008 Registered Company**

Annacotty Business Park, Annacotty, Limerick, Ireland.

Telephone 061 358190 Facsimile 061 358191