



'Ready To Use Reagents'

version 2011.01

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107
Lot Number 1121

Date 06.05.2011


Expiry Date 03.05.2012

Reagent	Lot Number
Phos 1	1120
Phos 2	1121
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 02 11	0.143
1	P20 02 11	0.290
1.5	P20 02 11	0.448
2	P20 02 11	0.577

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.001	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.291	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 107
* Based on results obtained in our Research Facility

Tom Russell.

Dr. T. Russell



An ISO 9001:2008 Registered Company

Annacotty Business Park, Annacotty, Limerick, Ireland.

Telephone 061 358190

Facsimile 061 358191