

Certificate of Performance
Sulphate Reagent


Product Code KR102
Lot Number 1164

Date 27.06.2011 **Expiry Date** 29.04.2012

Reagent	Lot Number
Sulphate	1164
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
250	Sulp 4000 02 11	0.091
500	Sulp 4000 02 11	0.217
800	Sulp 4000 02 11	0.389
1000	Sulp 4000 02 11	0.501

*serial Dilution

Calibration Curve


Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0004	0.0003 to 0.0006	Pass
250 - 500 mg/l	0.0005	0.0003 to 0.0007	Pass
500 - 800 mg/l	0.0006	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0006	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 102
 * Based on results obtained in our Research Facility



Dr. T. Russell