

### Certificate of Performance

## Phosphate 2 (Phos 2) Reagent

Product Code      KR107  
Lot Number            5004


Date            17.01.2005

Expiry Date      18.07.2005

<b>Reagent</b>	<b>Lot Number</b>
Phos 1	5003
Phos 2	5004
<b>Test Method</b>	LTP 05
<b>Test Instrument</b>	Konelab 20
<b>Serial Number</b>	P4919028

<b>Standard Value</b>	<b>Lot #</b>	<b>Abs.</b>
0		0.000
0.5	P0.5 349 04	0.150
1	P1.0 349 04	0.295
1.5	P1.5 349 04	0.441
2	P2.0 349 04	0.589

### Calibration Curve



Test		Result	Specification	Pass / Fail
Intercept		0.001	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm	0.294	0.260 to 0.320	Pass
Regression		1.000	0.990 to 1.001	Pass

*Tom Russell.*

Dr. T. Russell