

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code **KR107**
 Lot Number **5115**


Date 18.08.2005

Expiry Date 15.08.2006

Reagent	Lot Number
Phos 1	5114
Phos 2	5115
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P0.5 223 05	0.145
1	P1.0 223 05	0.288
1.5	P1.5 223 05	0.433
2	P2.0 223 05	0.577

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.001	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.288	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes
 Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 107
 * Based on results obtained in our Research Facility



Dr. T. Russell