

### Certificate of Performance

#### Phosphate 2 (Phos 2) Reagent

Product Code **KR107**  
 Lot Number **5123**


Date 16.09.2005

Expiry Date 15.09.2006

Reagent	Lot Number
Phos 1	5122
Phos 2	<b>5123</b>
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P0.5 223 05	0.144
1	P1.0 223 05	0.290
1.5	P1.5 223 05	0.432
2	P2.0 223 05	0.563

#### Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.282	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 107  
 \* Based on results obtained in our Research Facility



Dr. T. Russell