

Certificate of Performance

TON 2 Reagent

Product Code KR 104
Lot Number 5167


Date 06.12.2005

Expiry Date 05.12.2006

Reagent	Lot Number
TON 1	5166
TON 2	5167
TON 3	5168
Test Method	LMTPO3
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0		0.003
2.5	TON2.5 332 05	0.177
5	TON5.0 332 05	0.345
7.5	TON7.5 332 05	0.494
10	TON10 332 05	0.633

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.01	Pass
Sensitivity Abs/ppm			
0 - 2.5 mg/l	0.0071	0.0055 to 0.0080	Pass
2.5 - 5.0 mg/l	0.0067	0.0060 to 0.0080	Pass
5.0 - 7.5 mg/l	0.0060	0.0050 to 0.0075	Pass
7.5 - 10.0 mg/l	0.0056	0.0050 to 0.0070	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 104
* Based on results obtained in our Research Facility

Tom Russell.

Dr. T. Russell