

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107
Lot Number 6027


Date 20.01.2006

Expiry Date 19.01.2007

Reagent	Lot Number
Phos 1	5169
Phos 2	6027
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.000
0.5	P0.5 332 05	0.143
1	P1.0 332 05	0.284
1.5	P1.5 332 05	0.422
2	P2.0 332 05	0.560

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.280	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 107
* Based on results obtained in our Research Facility


Dr. T. Russell