

Certificate of Performance

Sulphate Reagent


Product Code KR102
Lot Number 6028

Date 25.01.2006 **Expiry Date** 19.01.2007

Reagent	Lot Number
Sulphate	6028
Test Method	LTP01 (1000)
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.003
250	SU250 332 05	0.087
500	SU500 332 05	0.217
750	SU750 332 05	0.361
1000	SU1000 332 05	0.504

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.010	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.00035	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.00052	0.0003 to 0.0006	Pass
500 - 750 mg/l	0.00058	0.0004 to 0.0007	Pass
750 - 1000 mg/l	0.00057	0.0004 to 0.0007	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 102
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell