

### Certificate of Performance

#### Phosphate 2 (Phos 2) Reagent

Product Code KR107  
 Lot Number 6040


Date 13.02.2006

Expiry Date 10.02.2007

Reagent	Lot Number
Phos 1	6039
Phos 2	6040
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.000
0.5	P0.5 332 05	0.145
1	P1.0 332 05	0.287
1.5	P1.5 332 05	0.434
2	P2.0 332 05	0.566

#### Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.284	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 107  
 \* Based on results obtained in our Research Facility

*Tom Russell*  
 Dr. T. Russell