

Certificate of Performance

Sulphate Reagent

Product Code KR102
Lot Number 6042


Date 20.02.2006

Expiry Date 10.02.2007

Reagent	Lot Number
Sulphate	6042
Test Method	LTP01 (1000)
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.003
250	SU250 332 05	0.088
500	SU500 332 05	0.216
750	SU750 332 05	0.353
1000	SU1000 332 05	0.497

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm		
	0 - 250 mg/l	0.00035 to 0.0005	Pass
	250 - 500 mg/l	0.00051 to 0.0006	Pass
	500 - 750 mg/l	0.00055 to 0.0007	Pass
	750 - 1000 mg/l	0.00058 to 0.0007	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 102
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell