

**Certificate of Performance**
**Nitrite Reagent**


**Product Code** KR105  
**Lot Number** 6095

**Date** 07.06.2006

**Expiry Date** 01.06.2007

Reagent	Lot Number
TON3	6095
Test Method	LTP04
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.000
0.5	NI0.5 332 05	0.411
1	NI1.0 332 05	0.839
1.5	NI1.5 332 05	1.267
2	NI2.0 332 05	1.680

**Calibration Curve**


Test	Result	Specification	Pass / Fail
Intercept	-0.004	-0.010 to 0.010	Pass
Sensitivity Abs/ppm	0.843	0.750 to 0.900	Pass
Regression	1.000	0.990 to 1.001	Pass

**Notes** Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 105  
 \* Based on results obtained in our Research Facility



Dr. T. Russell