

Certificate of Performance

Chloride Reagent


Product Code KR101
Lot Number 6116

Date 11.07.2006 **Expiry Date** 11.07.2007

Reagent	Lot Number
Chloride	6116
Test Method	LTP02
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Absorbance
0 ppm		0.003
250 ppm	CL25009006	0.315
500 ppm	CL50009006	0.521
750 ppm	CL75009006	0.670
1000 ppm	CL100009006	0.795

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.01	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0012	0.0010 to 0.0016	Pass
250 - 500 mg/l	0.0008	0.0006 to 0.0012	Pass
500 - 750 mg/l	0.0006	0.0004 to 0.0010	Pass
750 - 1000 mg/l	0.0005	0.0003 to 0.0008	Pass

Notes Store between 2 & 8 ° c
 *Stable on Board 5 days
 Barcode ID = 101
 * Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell