Precision®

Labmedics ANALYTICAL SOLUTIONS

'Ready To Use Reagents'

Ver 2006:02

Certificate of Performance

TON 2

Product Code KR104

Date 12.07.2006

Expiry Date 12.07.2007

Reagent	Lot Number
TON 1	6115
TON 2	**6117**
TON 3	6118
Test Method	LMTP03
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0		0.003
2.5	TON 100 187 06	0.178
5	TON 100 187 06	0.337
7.5	TON 100 187 06	0.498
10	TON 100 187 06	0.635

*serial dilution

Calibration Curve



Test		Result	Specification			Pass / Fail
Intercept		0.003	-0.005	to	0.01	**Pass**
Sensitivity	Abs/ppm					
	0 - 2.5 mg/l	0.0700	0.0550	to	0.0800	**Pass**
	2.5 - 5.0 mg/l	0.0636	0.0600	to	0.0800	**Pass**
	5.0 - 7.5 mg/l	0.0644	0.0600	to	0.0800	**Pass**
	7.5 - 10.0 mg/l	0.0548	0.0500	to	0.0700	**Pass**

Notes

Store between 2 & 8 °C

**Stable on Board TON 1 - 1 day*

** TON 2 & TON 3 - 5 days*

Keep Tightly Sealed when not in use.

Do not return unused portions to container

Barcode ID- TON 1 103; TON 2 104; TON 3 105

** Based on results obtained in our Research Facility*

Tom Russell

Dr. T. Russell

Labmedics ANALYTICAL SOLUTIONS

Telegraphic House, Waterfront Quay, Salford Quays, Manchester M50 3XW

Telephone 0161 869 0420