

Certificate of Performance

Ammonia 2 Reagent

Product Code **KR 109**
Lot Number **6120**

Date 13.07.06


Expiry Date 13.07.07

Reagent	Lot Number
AMM 1	6119
Amm 2	6120
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.027
0.25	amm 188 06	0.217
0.50	amm 188 06	0.421
0.77	amm 188 06	0.641
1.00	amm 188 06	0.830

**serial dilution*

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0212	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.8055	0.750 to 0.830	Pass
Regression	0.9998	0.990 to 1.0001	Pass

Notes *Store between 2 & 8 ° c*
**Stable on Board 5 days*
Barcode ID = 109
** Based on results obtained in our Research Facility*



Dr. T. Russell