

Certificate of Performance

Sulphate Reagent

Product Code KR102
Lot Number 6122


Date 18.07.2006 **Expiry Date** 13.06.2007

Reagent	Lot Number
Sulphate	6122
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.005
250	SU 250 332 05	0.106
500	SU 500 332 05	0.239
750	SU 750 332 05	0.395
1000	SU 1000 332 05	0.542

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.010	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0004	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.0005	0.0003 to 0.0006	Pass
500 - 750 mg/l	0.0006	0.0004 to 0.0007	Pass
750 - 1000 mg/l	0.0006	0.0004 to 0.0007	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 102
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell