

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106

Lot Number 6177

Date 01.10.2006


Expiry Date 01.10.2007

Reagent	Lot Number
Phos 1	6177
Phos 2	6176
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 08 06	0.145
1	P20 08 06	0.289
1.5	P20 08 06	0.441
2	P20 08 06	0.572

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.287	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility

Tom Russell
 Dr. T. Russell