

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107
Lot Number 6183

Date 18.10.06


Expiry Date 18.10.07

Reagent	Lot Number
Phos 1	6182
Phos 2	6183
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.003
0.5	P20 0806	0.149
1	P20 0806	0.297
1.5	P20 0806	0.454
2	P20 0806	0.586

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.294	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
*Stable on Board 5 days
Barcode ID = 106
* Based on results obtained in our Research Facility



Dr. T. Russell