

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107

Lot Number 7002

Date 4.01.2007


Expiry Date 4.01.2008

Reagent	Lot Number
Phos 1	7001
Phos 2	7002
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 08 06	0.148
1	P20 08 06	0.294
1.5	P20 08 06	0.449
2	P20 08 06	0.584

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.293	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility

Tom Russell
 Dr. T. Russell