

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106

Lot Number 7038

Date 15.02.2007


Expiry Date 15.02.2008

Reagent	Lot Number
Phos 1	7038
Phos 2	6108
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 08 06	0.145
1	P20 08 06	0.283
1.5	P20 08 06	0.442
2	P20 08 06	0.569

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.001	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.287	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility



Dr. T. Russell