

**Certificate of Performance**
**Phosphate 1 (Phos 1) Reagent**
**Product Code** KR106


**Lot Number** 7072

**Date** 03.04.2007

**Expiry Date** 03.04.2008

Reagent	Lot Number
Phos 1	7072
Phos 2	7073
<b>Test Method</b>	LTP 05
<b>Test Instrument</b>	Konelab 20
<b>Serial Number</b>	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 02 07	0.154
1	P20 02 07	0.293
1.5	P20 02 07	0.444
2	P20 02 07	0.582

*\*serial dilution*
**Calibration Curve**


	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.010	<b>Pass</b>
Sensitivity Abs/ppm	0.290	0.260 to 0.320	<b>Pass</b>
Regression	1.000	0.990 to 1.001	<b>Pass</b>

Notes Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 106  
 \* Based on results obtained in our Research Facility



Dr. T. Russell