

Certificate of Performance

**Phosphate 1 (Phos 1) Reagent**

**Product Code** KR106  
**Lot Number** 7087

**Date** 23.04.2007


**Expiry Date** 23.04.2008

Reagent	Lot Number
Phos 1	7087
Phos 2	7059
<b>Test Method</b>	LTP 05
<b>Test Instrument</b>	Konelab 20
<b>Serial Number</b>	P4919028

Standard Value	Lot #	Abs.
0		0.002
0.5	P20 02 07	0.141
1	P20 02 07	0.274
1.5	P20 02 07	0.419
2	P20 02 07	0.554

\*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.276	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

**Notes** Store between 2 & 8 °C  
 \*Stable on Board 5 days  
 Barcode ID = 106  
 \* Based on results obtained in our Research Facility

*Tom Russell.*  
 Dr. T. Russell