

Certificate of Performance
Ammonia 2 Reagent

Product Code KR 109
Lot Number 7152


Date 11.07.2007

Expiry Date 09.07.2008

Reagent	Lot Number
AMM 1	7151
Amm 2	7152
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.014
0.25	AMM 10 02 07	0.210
0.50	AMM 10 02 07	0.407
0.77	AMM 10 02 07	0.617
1.00	AMM 10 02 07	0.795

**serial dilution*

Calibration Curve


Test	Result	Specification	Pass / Fail
Intercept	0.0148	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.7814	0.750 to 0.890	Pass
Regression	1.0000	0.990 to 1.0001	Pass

Notes Store between 2 & 8 °C

*Stable on Board 5 days

Barcode ID = 109

* Based on results obtained in our Research Facility



Dr. T. Russell