

Certificate of Performance
TON 3

Product Code KR105


Date 03.08.2007

Expiry Date 30.08.2008

Reagent	Lot Number
TON 1	7160
TON 2	7161
TON 3	7162
Test Method	LMTPO3
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.004
2.50	TON 100 02 07	0.178
5.00	TON 100 02 07	0.336
7.69	TON 100 02 07	0.493
10.00	TON 100 02 07	0.626

*serial dilution

Calibration Curve


Test	Result	Specification	Pass / Fail
Intercept	0.004	-0.005 to 0.01	Pass
Sensitivity	Abs/ppm		
0 - 2.50 mg/l	0.0696	0.0550 to 0.0800	Pass
2.50 - 5.00 mg/l	0.0632	0.0600 to 0.0800	Pass
5.00 - 7.69 mg/l	0.0584	0.0550 to 0.0800	Pass
7.69 - 10.0 mg/l	0.0576	0.0500 to 0.0700	Pass

Notes Store between 2 & 8 °C
 *Stable on Board TON 1 - 1 day
 * TON 2 & TON 3 - 5 days
 Keep Tightly Sealed when not in use.
 Do not return unused portions to container
 Barcode ID- TON 1 103; TON 2 104; TON 3 105
 * Based on results obtained in our Research Facility



Dr. T. Russell