

Certificate of Performance

Ammonia 2 Reagent

Product Code **KR 109**
Lot Number **7166**

Date 30.07.2007


Expiry Date 30.07.2008

Reagent	Lot Number
AMM 1	7165
Amm 2	7166
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.017
0.25	Amm 10 02 07	0.224
0.50	Amm 10 02 07	0.434
0.77	Amm 10 02 07	0.655
1.00	Amm 10 02 07	0.853

**serial dilution*

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0161	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.8344	0.750 to 0.890	Pass
Regression	1.0000	0.990 to 1.0001	Pass

Notes *Store between 2 & 8 °C*
**Stable on Board 5 days*
Barcode ID = 109
** Based on results obtained in our Research Facility*



 Dr. T. Russell