

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107

Lot Number 7186

Date 30.08.2007


Expiry Date 27.08.2008

Reagent	Lot Number
Phos 1	7185
Phos 2	7186
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 08 07	0.155
1	P20 08 07	0.294
1.5	P20 08 07	0.441
2	P20 08 07	0.563

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.009	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.282	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 107
 * Based on results obtained in our Research Facility

Tom Russell.
 Dr. T. Russell