

Certificate of Performance
Phosphate 1 (Phos 1) Reagent
Product Code KR106


Lot Number 7222

Date 24.10.2007

Expiry Date 10.10.2008

Reagent	Lot Number
Phos 1	7222
Phos 2	7203
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 08 07	0.141
1	P20 08 07	0.276
1.5	P20 08 07	0.426
2	P20 08 07	0.551

**serial dilution*
Calibration Curve


	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.277	0.260 to 0.320	Pass
Regression	0.999	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C

**Stable on Board 5 days*

Barcode ID = 106

** Based on results obtained in our Research Facility*

Dr. T. Russell