Precision®

Labmedics ANALYTICAL SOLUTIONS

'Ready To Use Reagents' Ver 2006:03

Certificate of Performance

TON 2

Product Code KR104

Date 15.01.2008 **Expiry Date** 02.01.2009

Reagent	Lot Number
TON 1	8007
TON 2	**8001**
TON 3	8002
Test Method	LMTP03
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.009
2.50	TON 100 11 07	0.170
5.00	TON 100 11 07	0.350
7.69	TON 100 11 07	0.510
10.00	TON 100 11 07	0.670

*serial dilution

Calibration Curve



Test		Result	Specification			Pass / Fail
Intercept		0.009	-0.005	to	0.01	**Pass**
Sensitivity	Abs/ppm					
	0 - 2.50 mg/l	0.0644	0.0550	to	0.0800	**Pass**
	2.50 - 5.00 mg/l	0.0720	0.0600	to	0.0800	**Pass**
	5.00 - 7.69 mg/l	0.0595	0.0550	to	0.0800	**Pass**
	7.69 - 10.0 mg/l	0.0693	0.0500	to	0.0700	**Pass**

Notes *Store between 2 & 8 °C*

**Stable on Board TON 1 - 1 day*

** TON 2 & TON 3 - 5 days*

Keep Tightly Sealed when not in use.

Do not return unused portions to container

Barcode ID- TON 1 103; TON 2 104; TON 3 105

** Based on results obtained in our Research Facility*

Tom Russell.

Dr. T. Russell

Labmedics ANALYTICAL SOLUTIONS

Telegraphic House, Waterfront Quay, Salford Quays, Manchester M50 3XW

Telephone 0161 869 0420