

Certificate of Performance

Sulphate Reagent

Product Code KR102
Lot Number 8005

Date 10.01.2008


Expiry Date 02.01.2009

Reagent	Lot Number
Sulphate	8005
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.006
250	Sulp 4000 11 07	0.105
500	Sulp 4000 11 07	0.233
800	Sulp 4000 11 07	0.400
1000	Sulp 4000 11 07	0.495

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.006	-0.005 to 0.010	Pass
Sensitivity			
Abs/ppm			
0 - 250 mg/l	0.0004	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.0005	0.0003 to 0.0006	Pass
500 - 800 mg/l	0.0006	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0005	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 102
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell