Precision®

Labmedics ANALYTICAL SOLUTIONS

'Ready To Use Reagents' Ver 2006:03

Certificate of Performance

TON 3

Product Code KR105

Date 28.02.2008 **Expiry Date** 15.02.2009

Reagent	Lot Number
TON 1	8032
TON 2	8033
TON 3	**8034**
Test Method	LMTP03
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.003
2.50	TON 100 11 07	0.187
5.00	TON 100 11 07	0.357
7.69	TON 100 11 07	0.525
10.00	TON 100 11 07	0.648

*serial dilution

Calibration Curve



Test		Result	Specification			Pass / Fail
Intercept		0.003	-0.005	to	0.01	**Pass**
Sensitivity	Abs/ppm					
	0 - 2.50 mg/l	0.0736	0.0550	to	0.0800	**Pass**
	2.50 - 5.00 mg/l	0.0680	0.0600	to	0.0800	**Pass**
	5.00 - 7.69 mg/l	0.0625	0.0550	to	0.0800	**Pass**
	7.69 - 10.0 mg/l	0.0532	0.0500	to	0.0700	**Pass**

Notes *Store between 2 & 8 °C*
**Stable on Board TON 1 - 1 day*
** TON 2 & TON 3 - 5 days*
Keep Tightly Sealed when not in use.
Do not return unused portions to container
Barcode ID- TON 1 103; TON 2 104; TON 3 105
** Based on results obtained in our Research Facility*

Tom Russell.

Dr. T. Russell

Labmedics ANALYTICAL SOLUTIONS
Telegraphic House, Waterfront Quay, Salford Quays, Manchester M50 3XW