

Certificate of Performance
Phosphate 1 (Phos 1) Reagent
Product Code KR106


Lot Number 8114

Date 16.06.2008

Expiry Date 03.06.2009

Reagent	Lot Number
Phos 1	8114
Phos 2	8115
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.003
0.5	P20 02 08	0.149
1	P20 02 08	0.292
1.5	P20 02 08	0.451
2	P20 02 08	0.586

**serial dilution*
Calibration Curve


	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.294	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 * Based on results obtained in our Research Facility



Dr. T. Russell