

Certificate of Performance

Sulphate Reagent

Product Code KR102
Lot Number 8181

Date 15.09.2008


Expiry Date 11.09.2009

Reagent	Lot Number
Sulphate	8181
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.002
250	Sulp 4000 08 08	0.068
500	Sulp 4000 08 08	0.170
800	Sulp 4000 08 08	0.316
1000	Sulp 4000 08 08	0.420

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0003	0.0003 to 0.0005	Pass
250 - 500 mg/l	0.0004	0.0003 to 0.0006	Pass
500 - 800 mg/l	0.0005	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0005	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °c
*Stable on Board 5 days
Barcode ID = 102
* Based on results obtained in our Research Facility

Tom Russell

Dr. T. Russell