

Certificate of Performance

Ammonia 2 Reagent

Product Code KR 109
Lot Number 0068

Date 23.03.2010


Expiry Date 22.03.2011

Reagent	Lot Number
AMM 1	9121
Amm 2	0068
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.026
0.25	Amm 10 10 09	0.222
0.50	Amm 10 10 09	0.430
0.77	Amm 10 10 09	0.653
1.00	Amm 10 10 09	0.854

**serial dilution*

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0197	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.8280	0.750 to 0.890	Pass
Regression	0.9997	0.990 to 1.0001	Pass

Notes *Store between 2 & 8 °c*
 **Stable on Board 5 days*
 Barcode ID = 109
 ** Based on results obtained in our Research Facility*



 Dr. T. Russell