



Microgenics GmbH

Ready To Use Reagents'

ver 2008.03

Certificate of Performance

Sulphate Reagent

Product Code KR102

Lot Number 0069

Date 25.03.2010


Expiry Date 22.03.2011

Reagent	Lot Number
Sulphate	0069
Test Method	LTP01
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.006
250	Sulp 4000 11 09	0.128
500	Sulp 4000 11 09	0.276
800	Sulp 4000 11 09	0.451
1000	Sulp 4000 11 09	0.555

*serial Dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.006	-0.005 to 0.010	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0005	0.0003 to 0.0006	Pass
250 - 500 mg/l	0.0006	0.0003 to 0.0007	Pass
500 - 800 mg/l	0.0006	0.0004 to 0.0007	Pass
800 - 1000 mg/l	0.0005	0.0004 to 0.0008	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 102
 * Based on results obtained in our Research Facility

Dr. T. Russell