

Certificate of Performance

Chloride Reagent


Product Code KR101
Lot Number 0157

Date 30.07.2010 **Expiry Date** 27.07.2011

Reagent	Lot Number
Chloride	0157
Test Method	LTP02
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Absorbance
0 ppm		0.005
250 ppm	CL4000 02 10	0.312
500 ppm	CL4000 02 10	0.495
800 ppm	CL4000 02 10	0.645
1000 ppm	CL4000 02 10	0.723

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.005	-0.005 to 0.01	Pass
Sensitivity Abs/ppm			
0 - 250 mg/l	0.0012	0.0010 to 0.0016	Pass
250 - 500 mg/l	0.0008	0.0006 to 0.0012	Pass
500 - 800 mg/l	0.0005	0.0004 to 0.0010	Pass
800 - 1000 mg/l	0.0004	0.0003 to 0.0008	Pass

Notes Store between 2 & 8 °c
 *Stable on Board 5 days
 Barcode ID = 101
 * Based on results obtained in our Research Facility



Dr. T. Russell