



Microgenics GmbH

'Ready To Use Reagents'

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code KR107
Lot Number 0182

Date 27.09.2010


Expiry Date 23.09.2011

Reagent	Lot Number
Phos 1	0148
Phos 2	0182
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 09 10	0.147
1	P20 09 10	0.300
1.5	P20 09 10	0.440
2	P20 09 10	0.560

*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.007	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.282	0.260 to 0.320	Pass
Regression	0.998	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 107
 * Based on results obtained in our Research Facility

Tom Russell.

 Dr. T. Russell