Precision®

Microgenics GmbH *'Ready To Use Reagents'* Ver 2010:01

Certificate of Performance

TON 2

Product Code KR104

Date 13.12.2010 **Expiry Date** 01.12.2011

Reagent	Lot Number
TON 1	0265
TON 2	**0266**
TON 3	0163
Test Method	LMTP03
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.002
2.50	TON 100 09 10	0.180
5.00	TON 100 09 10	0.348
7.69	TON 100 09 10	0.507
10.00	TON 100 09 10	0.636

*serial dilution

Calibration Curve



Test		Result	Specification			Pass / Fail
Intercept		0.002	-0.005	to	0.01	**Pass**
Sensitivity	Abs/ppm					
	0 - 2.50 mg/l	0.0712	0.0550	to	0.0800	**Pass**
	2.50 - 5.00 mg/l	0.0672	0.0600	to	0.0800	**Pass**
	5.00 - 7.69 mg/l	0.0591	0.0550	to	0.0800	**Pass**
	7.69 - 10.0 mg/l	0.0558	0.0500	to	0.0700	**Pass**

Notes *Store between 2 & 8 °C*

**Stable on Board TON 1 - 1 day*

** TON 2 & TON 3 - 5 days*

Keep Tightly Sealed when not in use.

Do not return unused portions to container

Barcode ID- TON 1 103; TON 2 104; TON 3 105

** Based on results obtained in our Research Facility*

Tom Russell.

Dr. T. Russell